



**DEPARTMENT OF HEALTH CARE ACCESS AND INFORMATION  
FACILITIES DEVELOPMENT DIVISION**

**APPLICATION FOR HCAI SPECIAL SEISMIC  
CERTIFICATION PREAPPROVAL (OSP)**

OFFICE USE ONLY

**APPLICATION #: OSP-0280**

**HCAI Special Seismic Certification Preapproval (OSP)**

Type:  New  Renewal

**Manufacturer Information**

Manufacturer: Siemens Healthcare GmbH

Manufacturer's Technical Representative: Dieter Freitag

Mailing Address: Siemensstr. 3, D-91301 Forchheim, Germany

Telephone: +49 9191 185412

Email: freitag.dieter@siemens-healthineers.com

**Product Information**

Product Name: Fluoroscopy and Radiography Systems

Product Type: NA

Product Model Number: Ysio, Ysio Max, and Multix Fusion X-Ray Systems

General Description: Multi-component digital & analog radiographic medical imaging systems.

Mounting Description: Ceiling, Floor, Wall & floor, and Wall, See Certified Product Tables

Tested Seismic Enhancements: Seismic enhancements made to the test units and/or modifications required to address anomalies during the tests shall be incorporated into the production units.

**Applicant Information**

Applicant Company Name: W.E. Gundy & Associates, Inc.

Contact Person: Travis Soppe

Mailing Address: 1199 Shoreline Drive Suite 310, Boise, ID 83702

Telephone: (208) 342-5989

Email: tsoppe@wegai.com

Title: SE





**DEPARTMENT OF HEALTH CARE ACCESS AND INFORMATION  
FACILITIES DEVELOPMENT DIVISION**

**California Licensed Structural Engineer Responsible for the Engineering and Test Report(s)**

Company Name: W.E. GUNDY & ASSOCIATES INC.  
Name: Travis Soppe California License Number: S6115  
Mailing Address: P.O. Box 9121, Boise, ID 83707  
Telephone: (208) 342-5989 Email: tsoppe@wegai.com

**Certification Method**

GR-63-Core       ICC-ES AC156       IEEE 344       IEEE 693       NEBS 3  
 Other (Please Specify): \_\_\_\_\_

**Testing Laboratory**

Company Name: ENVIRONMENTAL TESTING LABORATORIES, INC. (ETL)  
Contact Person: Brady Richard  
Mailing Address: 11034 Indian Trail, Dallas TX 75229-3513  
Telephone: (972) 247-9657 Email: brady@etldallas.com





**DEPARTMENT OF HEALTH CARE ACCESS AND INFORMATION  
FACILITIES DEVELOPMENT DIVISION**

**Seismic Parameters**

Design Basis of Equipment or Components ( $F_p/W_p$ ) = See attachments

SDS (Design spectral response acceleration at short period, g) = 2.0

$a_p$  (Amplification factor) = See attachments

$R_p$  (Response modification factor) = See attachments

$\Omega_0$  (System overstrength factor) = See Attachment

$I_p$  (Importance factor) = 1.5

$z/h$  (Height ratio factor) = 1

Natural frequencies (Hz) = See Attachment

Overall dimensions and weight = See Attachment

**HCAI Approval (For Office Use Only) - Approval Expires on 11/30/2028**

Date: 11/30/2022

Name: Mohammad Karim Title: Supervisor, Health Facilities

Special Seismic Certification Valid Up to: SDS (g) = 2.0 z/h = 1

Condition of Approval (if applicable): DATE: 11/30/2022



<b>TABLE 1</b>	<b>SIEMENS HEALTHCARE GmbH</b>					 <b>WEGAI</b> <small>W.E. GUNDY &amp; ASSOCIATES, INC.</small> <small>STRUCTURAL &amp; EARTHQUAKE ENGINEERING</small>		
	<b>SPECIAL SEISMIC CERTIFICATION</b>							
<b>YSIO &amp; YSIO MAX COMPONENTS</b>								

**Manufacturer:** Siemens Healthcare GmbH

**System:** YSIO & YSIO MAX

System Component <sup>1)</sup>	Siemens Part Number	Dimensions (in)			Weight (lb)	Mounting	UUT
		Width	Depth	Height			

**Ceiling Suspension Ysio X-Ray Tubes**

3D Stand Ysio 3m track-synchronized	7042232	167	119	32 - 103	754	ceiling	UUT <sub>x</sub> -1
3D Stand Ysio 4m track-synchronized	7042240	167	172	32 - 103	815	ceiling	interpolated
3D Stand Ysio 3m track-automated	7042216	167	119	32 - 103	846	ceiling	interpolated
3D Stand Ysio 4m track-automated	7042224	167	172	32 - 103	890	ceiling	UUT <sub>w</sub> -1

**Bucky Wall Stands**

Pro Non-Tilting Bucky Wall Stand - Digital Detector	10681650	30.0	16.6	82.9	452	floor	UUT <sub>v</sub> -7
Non-Tilting Bucky Wall Stand - Fixed Detector	11688339	30.0	18.5	82.9	488	floor	interpolated
Non-Tilting Bucky Wall Stand - Wifi Dectector	11688339	30.0	18.5	82.9	488	floor	interpolated
Tilting Bucky Wall Stand Fixed Detector	10681702	30.0	48.9	85.0	530	floor	UUT <sub>y</sub> -2
Tilting Bucky Wall Stand Wifi Dectector	10681704	30.0	28.1	82.9	645	floor	UUT <sub>z</sub> -2
Tilting Bucky Wall Stand Wifi Dectector	10681705	30.0	28.1	82.9	645	floor	interpolated
Tilting Bucky Wall Stand Wifi Dectector	10656443	30.0	26.9	82.5	660	floor	interpolated
Tilting Bucky Wall Stand Wifi Dectector	10656442	30.0	26.9	82.5	660	floor	interpolated
Tilting Bucky Wall Stand Wifi Dectector	10150520	30.0	26.9	82.5	660	floor	interpolated
Tilting Bucky Wall Stand Wifi Dectector	10150519	30.0	26.9	82.5	660	floor	UUT <sub>w</sub> -4

<sup>1</sup> All components are manufactured by Siemens Healthcare GmbH unless noted. Part numbers listed uniquely identify type of component, manufacturer, and material of construction for each sub-component within the tested units.

<sup>2</sup> The units were tested at different times and the subscripts on the UUTs reference the following seismic test reports:  
 $v = \text{SQ35-1302-02-r1}$     $w = \text{SQ35-1204-01-r3}$     $x = \text{SSC10-1010-02-r2}$     $y = \text{SQ35-1416-01-r2}$     $z = \text{SQ35-1415-02-r4}$

**SEISMIC CERTIFICATION LIMITS**

System Component	Code	S <sub>DS</sub> (g)	z / h	I <sub>p</sub>	a <sub>p</sub>	R <sub>p</sub>	Ω <sub>0</sub>	F <sub>p</sub> / W <sub>p</sub>
Ceiling Suspension X-Ray Tubes	CBC 2022	2.0	1.0	1.50	2.5	2.5	2.0	3.60
Wall Stands	CBC 2022	2.0	1.0	1.50	1.0	1.5	1.5	2.40

<b>TABLE 1</b>	<b>SIEMENS HEALTHCARE GmbH SPECIAL SEISMIC CERTIFICATION YSIO &amp; YSIO MAX COMPONENTS</b>					 <b>WEGAI</b> <small>W.E. GUNDY &amp; ASSOCIATES, INC. STRUCTURAL &amp; EARTHQUAKE ENGINEERING</small>		
	<b>Manufacturer:</b> Siemens Healthcare GmbH							

**System:** YSIO & YSIO MAX

System Component <sup>1)</sup>	Siemens Part Number	Dimensions (in)			Weight (lb)	Mounting	UUT
		Width	Depth	Height			

**Patient Tables**

Ysio MAX Patient Table	10273210	94.9	31.5	20.3 - 37.6	835 <sup>3)</sup>	floor	UUT <sub>y</sub> -1
Ysio Patient Table	10281013	94.9	31.5	20.3 - 37.5	1015 <sup>3)</sup>	floor	UUT <sub>w</sub> -2

**Imaging Systems**

Flurospot Compact (FLC)	10762484	13.4	25.9	21.8	85	floor	UUT <sub>y</sub> -3
Flurospot Compact (FLC)	10281016	13.4	27.0	21.8	116	floor	UUT <sub>w</sub> -6

**Generator Cabinets**

Polydoros 80kW	10096950	31.5	17.1	86.7	835	floor/wall	UUT <sub>y</sub> -6
Polydoros R80/R65	10096980	31.0	17.0	86.7	883	floor/wall	UUT <sub>w</sub> -5

**Miscellaneous Components: Docking Station and Wireless Access Point**

Docking Station	10281025	20.5	8.0	18.0	33	wall	UUT <sub>w</sub> -3
SCALANCE W700	Scalance W700	7.9	2.8	6.2	6	wall	UUT <sub>y</sub> -4

<sup>1</sup> All components are manufactured by Siemens Healthcare GmbH unless noted. Part numbers listed uniquely identify type of component, manufacturer, and material of construction for each sub-component within the tested units.  
<sup>2</sup> The units were tested at different times and the subscripts on the UUTs reference the following seismic test reports:  
w = SQ35-1204-01-r3    x = SSC10-1010-02-r2    y = SQ35-1416-01-r2    z = SQ35-1415-02-r4  
<sup>3</sup> Patient table weight does not include simulated patient weights of 440lb and 660lb for UUT<sub>y</sub>-1 and UUT<sub>w</sub>-2, respectively.

**SEISMIC CERTIFICATION LIMITS**

System Component	Code	S <sub>DS</sub> (g)	z / h	I <sub>P</sub>	a <sub>P</sub>	R <sub>P</sub>	Ω <sub>0</sub>	F <sub>P</sub> / W <sub>P</sub>
Patient Tables	CBC 2022	2.0	1.0	1.50	1.0	1.5	1.5	2.40
Imaging Systems	CBC 2022	2.0	1.0	1.50	1.0	2.5	2.0	1.44
Generator Cabinets	CBC 2022	2.0	1.0	1.50	2.5	6.0	2.0	1.50
Docking Station / Wireless Access Point	CBC 2022	2.0	1.0	1.50	1.0	2.5	2.0	1.44

<b>TABLE 1</b>	<b>SIEMENS HEALTHCARE GmbH SPECIAL SEISMIC CERTIFICATION MULTIX FUSION COMPONENTS</b>					 <b>WEGAI</b> <small>W.E. GUNDY &amp; ASSOCIATES, INC. STRUCTURAL &amp; EARTHQUAKE ENGINEERING</small>	

**Manufacturer:** Siemens Healthcare GmbH

**System:** Multix Fusion

System Component <sup>1)</sup>	Siemens Part Number	Dimensions (in)			Weight (lb)	Mounting	UUT
		Width	Depth	Height			

### Ceiling Suspension X-Ray Tubes

3D V Stand ACCS T 3m track-synchronized	7042166	167	119	32 - 103	750	ceiling	UUT <sub>v</sub> -6
3D V Stand ACCS T 4m track-synchronized	7042174	167	119	32 - 103	815	ceiling	interpolated
3D Stand Ysio 4m track-automated	7042224	167	172	32 - 103	890	ceiling	UUT <sub>w</sub> -1

### Wall Stands

Top Tilting Wall Stand	10681670	30.0	28.1	82.9	490	floor	UUT <sub>v</sub> -8
Pro Non-Tilting Wall Stand	10681650	30.0	16.6	82.9	452	floor	UUT <sub>v</sub> -7

### Patient Tables

Multix Fusion Patient Table	10273204	94.8	31.5	20.3 - 37.6	886 <sup>2)</sup>	floor	UUT <sub>v</sub> -9
--------------------------------	----------	------	------	-------------	-------------------	-------	---------------------

<sup>1</sup> All components are manufactured by Siemens Healthcare GmbH unless noted. Part numbers listed uniquely identify type of component, manufacturer, and material of construction for each sub-component within the tested units.

<sup>2</sup> Patient table weight does not include 660lb simulated patient weight.

<sup>3</sup> All wall stands exclude the optional Spacer Box.

<sup>4</sup> Wall stands were modified as detailed in the test report, SQ35-1202-02 Rev. 1, these enhancements are required to meet the conditions of this listing.

<sup>5</sup> The units were tested at different times and the subscripts on the UUTs reference the following seismic test reports:

v = SQ35-1302-02-r1      w = SQ35-1204-01-r3

### SEISMIC CERTIFICATION LIMITS

System Component	Code	S <sub>DS</sub> (g)	z / h	I <sub>P</sub>	a <sub>P</sub>	R <sub>P</sub>	Ω <sub>0</sub>	F <sub>P</sub> / W <sub>P</sub>
Ceiling Suspension X-Ray Tubes	CBC 2022	2.0	1.0	1.50	2.5	2.5	2.0	3.60
Wall Stands	CBC 2022	2.0	1.0	1.50	1.0	1.5	1.5	2.40
Patient Tables	CBC 2022	2.0	1.0	1.50	1.0	1.5	1.5	2.40

<b>TABLE 1</b>	<b>SIEMENS HEALTHCARE GmbH SPECIAL SEISMIC CERTIFICATION MULTIX FUSION COMPONENTS</b>	 <b>WEGAI</b> W.E. GUNDY & ASSOCIATES, INC. STRUCTURAL & EARTHQUAKE ENGINEERING
----------------	---	---

**Manufacturer:** Siemens Healthcare GmbH

**System:** Multix Fusion

System Component <sup>1)</sup>	Siemens Part Number	Dimensions (in)			Weight (lb)	Mounting	UUT
		Width	Depth	Height			
<b>Imaging Systems</b>							
Flurospot Compact	10569621	17.6	27.6	22.8	158	floor	UUT <sub>v</sub> -3
<b>Generator Cabinets</b>							
Polydoros RF RAD80 / 65 / 55	10606040	22.5	50.0	21.4	424	floor	UUT <sub>v</sub> -1
<b>Miscellaneous Components: Power Box</b>							
Power Box	10847706	12.4	2.3	11.0	10	wall	UUT <sub>v</sub> -2

<sup>1</sup> All components are manufactured by Siemens Healthcare GmbH unless noted. Part numbers listed uniquely identify type of component, manufacturer, and material of construction for each sub-component within the tested units.

<sup>2</sup> Patient table weight does not include 660lb simulated patient weight.

<sup>3</sup> All wall stands exclude the optional Spacer Box.

<sup>4</sup> Wall stands were modified as detailed in the test report, SQ35-1202-02 Rev. 1, these enhancements are required to meet the conditions of this listing.

<sup>5</sup> The units were tested at different times and the subscripts on the UUTs reference the following seismic test reports:

v = SQ35-1302-02-r1      w = SQ35-1204-01-r3

**SEISMIC CERTIFICATION LIMITS**

System Component	Code	S <sub>DS</sub> (g)	z / h	I <sub>p</sub>	a <sub>p</sub>	R <sub>p</sub>	Ω <sub>0</sub>	F <sub>p</sub> / W <sub>p</sub>
Imaging Systems	CBC 2022	2.0	1.0	1.50	1.0	2.5	2.0	1.44
Generator Cabinets	CBC 2022	2.0	1.0	1.50	2.5	6.0	2.0	1.50
Power Box	CBC 2022	2.0	1.0	1.50	1.0	2.5	2.0	1.44

UUT<sub>w</sub>-1

**UNIT UNDER TEST (UUT)  
SUMMARY SHEET**



**Mounting Details:** Rails and connecting parts of the component bolt with 2 -M10 bolts (20 bolts total) to unistrut grid spaced at 27.0" on center. The unistrut grid consisted of Unistrut P1001 rails anchored with 2 - 3/8" bolts at each intersection to the ceiling fixture framing spaced at 32" on center.



<b>Manufacturer:</b> Siemens Healthcare GmbH	<b>Test Location:</b> Environmental Testing Laboratory
<b>Component:</b> X-Ray Tube Stand w/ 4m Bridge	<b>Test Date:</b> March / May 2012
<b>Model Number:</b> 7042224	<b>Report Number:</b> SQ35-1204-01 Rev. 3
<b>UUT Function:</b> X-Ray stand, ceiling suspended, for use in radiography imaging, fully automated	
<b>UUT Description:</b> Component of the YSIO and YSIO MAX Systems	

**UUT PROPERTIES**

Weight (lb)	Dimensions (inches)			Natural Frequency (Hz)		
	Width	Depth	Height	FB	SS	V
888	167	172	32-103	N/A	N/A	N/A

The ceiling suspended tube stand moves laterally, rotates, and extends up and down to accommodate different patients and procedures. The system was tested in the normal operating position with the system horizontally centered, no rotation, and with a height of 50" from mounting point to bottom of collimator.

**SEISMIC TEST PARAMETERS**

Building Code / Test Criteria	S <sub>DS</sub> (g)	z / h	I <sub>p</sub>	A <sub>FLX-H</sub> (g)	A <sub>RIG-H</sub> (g)	A <sub>FLX-V</sub> (g)	A <sub>RIG-V</sub> (g)
CBC 2022 / ICC-ES AC156	2.00	1.0	1.5	3.20	2.40	1.34	0.54

Note: The unit was full of contents during testing and remained functional before and after the ICC-ES AC156 test. The unit maintained structural integrity during and after the ICC-ES AC156 Test.

UUT<sub>w</sub>-2

**UNIT UNDER TEST (UUT)  
SUMMARY SHEET**



**Mounting Details:** Rigid floor mounted with 4 - 5/8" bolts



**Manufacturer:** Siemens Healthcare GmbH | **Test Location:** Environmental Testing Laboratory

**Component:** Patient Table | **Test Date:** March 2012

**Model Number:** 10281013 | **Report Number:** SQ35-1204-01 Rev. 3

**UUT Function:** Motorized patient table for support and positioning for image acquisition.

**UUT Description:** Component of the YSIO and YSIO MAX systems, includes wi-D Digital Deceptor - Trixell Pixium FE 3542 pR (P/N: 62155967)

**UUT PROPERTIES**

Weight w/ Patient(lb)	Dimensions (inches)			Natural Frequency (Hz)		
	Width	Depth	Height	FB	SS	V
1,675	94.9	31.5	20.3-37.5	2.4	23.0	11.7

The patient table moves laterally both ways and vertically to accommodate different patients and procedures. The system was tested in the normal operating position with the table horizontally centered, a table top height of 31.3", and a total simulated patient weight of 660lbs.

**SEISMIC TEST PARAMETERS**

Building Code / Test Criteria	S <sub>DS</sub> (g)	z / h	I <sub>p</sub>	A <sub>FLX-H</sub> (g)	A <sub>RIG-H</sub> (g)	A <sub>FLX-V</sub> (g)	A <sub>RIG-V</sub> (g)
CBC 2022 / ICC-ES AC156	2.00	1.0	1.5	3.20	2.40	1.34	0.54

Note: The unit was full of contents during testing and remained functional before and after the ICC-ES AC156 test. The unit maintained structural integrity during and after the ICC-ES AC156 Test.

UUT<sub>w</sub>-3

**UNIT UNDER TEST (UUT)  
SUMMARY SHEET**



**Mounting Details:** Rigid wall mounted with 4 - 1/4" sheet metal screws



**Manufacturer:** Siemens Healthcare GmbH | **Test Location:** Environmental Testing Laboratory

**Component:** Docking Station w/ WI-D Detector | **Test Date:** March 2012

**Model Number:** 10281025 | **Report Number:** SQ35-1204-01 Rev. 3

**UUT Function:** Wall mounted docking station for detector used in radiographic medical imaging

**UUT Description:** Component of the YSIO and YSIO MAX systems includes wi-D digital detector, Trixell Pixium FE 3542 pR (P/N: 62155967)

**UUT PROPERTIES**

Weight (lb)	Dimensions (inches)			Natural Frequency (Hz)		
	Width	Depth	Height	FB	SS	V
33	20.5	8	18	N/A	N/A	N/A

**SEISMIC TEST PARAMETERS**

Building Code / Test Criteria	S <sub>DS</sub> (g)	z / h	I <sub>p</sub>	A <sub>FLX-H</sub> (g)	A <sub>RIG-H</sub> (g)	A <sub>FLX-V</sub> (g)	A <sub>RIG-V</sub> (g)
CBC 2022 / ICC-ES AC156	2.00	1.0	1.5	3.20	2.40	1.34	0.54

Note: The unit was full of contents during testing and remained functional before and after the ICC-ES AC156 test. The unit maintained structural integrity during and after the ICC-ES AC156 Test.

UUT<sub>w</sub>-4

**UNIT UNDER TEST (UUT)  
SUMMARY SHEET**



**Mounting Details:** Rigid floor mounted with 4 - 5/8" bolts



**Manufacturer:** Siemens Healthcare GmbH **Test Location:** Environmental Testing Laboratory

**Component:** Wall Stand **Test Date:** March 2012

**Model Number:** 10150519 **Report Number:** SQ35-1204-01 Rev. 3

**UUT Function:** Wall stand for radiographic medical imaging

**UUT Description:** Component of the YSIO and YSIO MAX systems includes wi-D Digital Detector, Trixell Pixium FE 3542 pR (P/N: 62155967)

**UUT PROPERTIES**

Weight (lb)	Dimensions (inches)			Natural Frequency (Hz)		
	Width	Depth	Height	FB	SS	V
661	30	26.9	82.5	8.8	8.1	8.8

**SEISMIC TEST PARAMETERS**

Building Code / Test Criteria	S <sub>DS</sub> (g)	z / h	I <sub>p</sub>	A <sub>FLX-H</sub> (g)	A <sub>RIG-H</sub> (g)	A <sub>FLX-V</sub> (g)	A <sub>RIG-V</sub> (g)
CBC 2022 / ICC-ES AC156	2.00	1.0	1.5	3.20	2.40	1.34	0.54

Note: The unit was full of contents during testing and remained functional before and after the ICC-ES AC156 test. The unit maintained structural integrity during and after the ICC-ES AC156 Test.

UUT<sub>w</sub>-5

**UNIT UNDER TEST (UUT)  
SUMMARY SHEET**



**Mounting Details:** Rigid wall mounted at top back side of unit with 4 - 1/4" grade 8 bolts and rigid floor mounted with 2 - 3/8" grade 8 bolts. Wall mount brackets constructed of 4" long L4x2.5x0.25 angles that bolt thru the short leg to the UUT with a single 1/2" grade 8 bolt.



**Manufacturer:** Siemens Healthcare GmbH | **Test Location:** Environmental Testing Laboratory

**Component:** Polydoros R80 Generator Cabinet | **Test Date:** March 2012

**Model Number:** 10096980 | **Report Number:** SQ35-1204-01 Rev. 3

**UUT Function:** Generator for radiography and fluoroscopy systems

**UUT Description:** Component of the YSIO and YSIO MAX systems

**UUT PROPERTIES**

Weight (lb)	Dimensions (inches)			Natural Frequency (Hz)		
	Width	Depth	Height	FB	SS	V
883	31	17	86.8	N/A	N/A	N/A

**SEISMIC TEST PARAMETERS**

Building Code / Test Criteria	S <sub>DS</sub> (g)	z / h	I <sub>p</sub>	A <sub>FLX-H</sub> (g)	A <sub>RIG-H</sub> (g)	A <sub>FLX-V</sub> (g)	A <sub>RIG-V</sub> (g)
CBC 2022 / ICC-ES AC156	2.00	1.0	1.5	3.20	2.40	1.34	0.54

Note: The unit was full of contents during testing and remained functional before and after the ICC-ES AC156 test. The unit maintained structural integrity during and after the ICC-ES AC156 Test.

UUT<sub>w</sub>-6

**UNIT UNDER TEST (UUT)  
SUMMARY SHEET**



**Mounting Details:** Rigid floor mounted with 4 - 3/8" cap screws with washers



**Manufacturer:** Siemens Healthcare GmbH | **Test Location:** Environmental Testing Laboratory

**Component:** Flurospot Compact Imaging System | **Test Date:** March 2012

**Model Number:** 10281016 | **Report Number:** SQ35-1204-01 Rev. 3

**UUT Function:** Digital imaging solution for fluoroscopy and radiography

**UUT Description:** Component of the YSIO and YSIO MAX systems

**UUT PROPERTIES**

Weight (lb)	Dimensions (inches)			Natural Frequency (Hz)		
	Width	Depth	Height	FB	SS	V
116	13.4	27	21.9	21.6	6.7	>33

**SEISMIC TEST PARAMETERS**

Building Code / Test Criteria	S <sub>DS</sub> (g)	z / h	I <sub>p</sub>	A <sub>FLX-H</sub> (g)	A <sub>RIG-H</sub> (g)	A <sub>FLX-V</sub> (g)	A <sub>RIG-V</sub> (g)
CBC 2022 / ICC-ES AC156	2.00	1.0	1.5	3.20	2.40	1.34	0.54

Note: The unit was full of contents during testing and remained functional before and after the ICC-ES AC156 test. The unit maintained structural integrity during and after the ICC-ES AC156 Test.

UUT<sub>y</sub>-1

**UNIT UNDER TEST (UUT)  
SUMMARY SHEET**



**Mounting Details:** Rigid floor mounted using 4 - 5/8" grade 8 bolts with washers



OSP-0280

BY: Mohammad Karim

**Manufacturer:** Siemens Healthcare GmbH | **Test Location:** Environmental Testing Laboratory

**Component:** YSIO MAX Patient Table | **Test Date:** October 2014

**Model Number:** 10273210 | **Report Number:** SQ35-1416-01 Rev. 2

**UUT Function:** Motorized patient table for support and positioning for image acquisition

**UUT Description:** Component of YSIO and YSIO MAX systems, includes MAX wi-D wireless mobile digital detector - Trixell Pixium 3543EZh (P/N: 10762402)

**UUT PROPERTIES**

Weight w/ Patient(lb)	Dimensions (inches)			Natural Frequency (Hz)		
	Width	Depth	Height	FB	SS	V
1,275	94.9	31.5	20.3 - 37.6	25.2	>33	11.6

The patient table moves laterally both ways and vertically to accommodate different patients and procedures. The system was tested in the normal operating position with the table horizontally centered, a table top height of 30", and a total simulated patient weight of 440lbs.

**SEISMIC TEST PARAMETERS**

Building Code / Test Criteria	S <sub>DS</sub> (g)	z / h	I <sub>p</sub>	A <sub>FLX-H</sub> (g)	A <sub>RIG-H</sub> (g)	A <sub>FLX-V</sub> (g)	A <sub>RIG-V</sub> (g)
CBC 2022 / ICC-ES AC156	2.00	1.0	1.5	3.20	2.40	1.34	0.54

Note: The unit was full of contents during testing and remained functional before and after the ICC-ES AC156 test. The unit maintained structural integrity during and after the ICC-ES AC156 Test.

UUT<sub>y</sub>-2

**UNIT UNDER TEST (UUT)  
SUMMARY SHEET**



**Mounting Details:** Rigid floor mounted with 4 - 5/8" grade 8 bolts



<b>Manufacturer:</b> Siemens Healthcare GmbH	<b>Test Location:</b> Environmental Testing Laboratory
<b>Component:</b> Bucky Wall Stand w/ MAX Static	<b>Test Date:</b> October 2014
<b>Model Number:</b> 10681702	<b>Report Number:</b> SQ35-1416-01 Rev. 2
<b>UUT Function:</b> Radiographic wall stand for X-ray exposures	
<b>UUT Description:</b> Component of YSIO and YSIO MAX systems includes Trixell MAX Static fixed plate detector (P/N: 10762401)	

**UUT PROPERTIES**

Weight (lb)	Dimensions (inches)			Natural Frequency (Hz)		
	Width	Depth	Height	FB	SS	V
528	30	48.9	85	8.6	7.7	8.7

**SEISMIC TEST PARAMETERS**

Building Code / Test Criteria	S <sub>DS</sub> (g)	z / h	I <sub>p</sub>	A <sub>FLX-H</sub> (g)	A <sub>RIG-H</sub> (g)	A <sub>FLX-V</sub> (g)	A <sub>RIG-V</sub> (g)
CBC 2022 / ICC-ES AC156	2.00	1.0	1.5	3.20	2.40	1.34	0.54

Note: The unit was full of contents during testing and remained functional before and after the ICC-ES AC156 test. The unit maintained structural integrity during and after the ICC-ES AC156 Test.

UUT<sub>y</sub>-3

**UNIT UNDER TEST (UUT)  
SUMMARY SHEET**



**Mounting Details:** Rigid floor mounted with 4 - 3/8" bolts



BY: Mohammad Karim

DATE: 11/30/2022

**Manufacturer:** Siemens Healthcare GmbH | **Test Location:** Environmental Testing Laboratory

**Component:** Fluorospot Compact (FLC) | **Test Date:** October 2014

**Model Number:** 10762484 | **Report Number:** SQ35-1416-01 Rev. 2

**UUT Function:** Digital imaging solution for fluoroscopy and radiography

**UUT Description:** Component of the YSIO and YSIO MAX systems

**UUT PROPERTIES**

Weight (lb)	Dimensions (inches)			Natural Frequency (Hz)		
	Width	Depth	Height	FB	SS	V
85	13.4	26	21.7	20.7	12.7	26.3

**SEISMIC TEST PARAMETERS**

Building Code / Test Criteria	S <sub>DS</sub> (g)	z / h	I <sub>P</sub>	A <sub>FLX-H</sub> (g)	A <sub>RIG-H</sub> (g)	A <sub>FLX-V</sub> (g)	A <sub>RIG-V</sub> (g)
CBC 2022 / ICC-ES AC156	2.00	1.0	1.5	3.20	2.40	1.34	0.54

Note: The unit was full of contents during testing and remained functional before and after the ICC-ES AC156 test. The unit maintained structural integrity during and after the ICC-ES AC156 Test.

UUT<sub>y</sub>-4

**UNIT UNDER TEST (UUT)  
SUMMARY SHEET**



**Mounting Details:** Rigid wall mounted with 4 - 1/4" screws



**Manufacturer:** Siemens Healthcare GmbH **Test Location:** Environmental Testing Laboratory

**Component:** Wireless access point **Test Date:** October 2014

**Model Number:** Scalance W700 **Report Number:** SQ35-1416-01 Rev. 2

**UUT Function:** Wireless access point

**UUT Description:** Component of the YSIO and YSIO MAX systems

**UUT PROPERTIES**

Weight (lb)	Dimensions (inches)			Natural Frequency (Hz)		
	Width	Depth	Height	FB	SS	V
6	7.9	2.8	6.2	N/A	N/A	N/A

**SEISMIC TEST PARAMETERS**

Building Code / Test Criteria	S <sub>DS</sub> (g)	z / h	I <sub>p</sub>	A <sub>FLX-H</sub> (g)	A <sub>RIG-H</sub> (g)	A <sub>FLX-V</sub> (g)	A <sub>RIG-V</sub> (g)
CBC 2022 / ICC-ES AC156	2.00	1.0	1.5	3.20	2.40	1.34	0.54

Note: The unit was full of contents during testing and remained functional before and after the ICC-ES AC156 test. The unit maintained structural integrity during and after the ICC-ES AC156 Test.

UUT<sub>y</sub>-6

**UNIT UNDER TEST (UUT)  
SUMMARY SHEET**



**Mounting Details:** Rigid wall mounted at top back side of unit with 4 - 1/4" grade 8 bolts and rigid floor mounted with 2 - 5/8" grade 8 bolts. Wall mount brackets constructed of 4" long L4x2.5x0.25 angles that bolt thru the short leg to the UUT with a single 1/2" grade 8 bolt.



REVIEWED FOR CODE COMPLIANCE  
HCAI  
OSP-0280  
BY: Mohammad Karim  
DATE: 11/30/2022  
CALIFORNIA BUILDING CODE, 2019

**Manufacturer:** Siemens Healthcare GmbH | **Test Location:** Environmental Testing Laboratory

**Component:** Polydoros Generator Cabinet | **Test Date:** October 2014

**Model Number:** 10096950 | **Report Number:** SQ35-1416-01 Rev. 2

**UUT Function:** Generator for radiography and fluoroscopy systems

**UUT Description:** Component of the YSIO and YSIO MAX systems

**UUT PROPERTIES**

Weight (lb)	Dimensions (inches)			Natural Frequency (Hz)		
	Width	Depth	Height	FB	SS	V
835	31.5	17.1	86.7	N/A	N/A	N/A

**SEISMIC TEST PARAMETERS**

Building Code / Test Criteria	S <sub>DS</sub> (g)	z / h	I <sub>p</sub>	A <sub>FLX-H</sub> (g)	A <sub>RIG-H</sub> (g)	A <sub>FLX-V</sub> (g)	A <sub>RIG-V</sub> (g)
CBC 2022 / ICC-ES AC156	2.00	1.0	1.5	3.20	2.40	1.34	0.54

Note: The unit was full of contents during testing and remained functional before and after the ICC-ES AC156 test. The unit maintained structural integrity during and after the ICC-ES AC156 Test.

UUT<sub>x</sub>-1

**UNIT UNDER TEST (UUT)  
SUMMARY SHEET**



**Mounting Details:** Rails and connecting parts of the component bolt with 2 -M10 bolts (20 bolts total) to unistrut grid spaced at 26.0" on center. The unistrut grid consisted of Unistrut P1001 rails anchored with 2 - 3/8" bolts at each intersection to the ceiling fixture framing spaced at 20" on center.



<b>Manufacturer:</b> Siemens Healthcare GmbH	<b>Test Location:</b> Environmental Testing Laboratory
<b>Component:</b> X-Ray Tube Stand w/ 3m Bridge	<b>Test Date:</b> April 2010
<b>Model Number:</b> 7042232	<b>Report Number:</b> SSC10-1010-2 Rev. 2
<b>UUT Function:</b> X-Ray stand, ceiling suspended, for use in radiography imaging, fully synchronized	
<b>UUT Description:</b> Component of the Luminos dRF System	

**UUT PROPERTIES**

Weight (lb)	Dimensions (inches)			Natural Frequency (Hz)		
	Width	Depth	Height	FB	SS	V
754	120	126	106	N/A	N/A	N/A

The ceiling suspended tube stand moves laterally, rotates, and extends up and down to accommodate different patients and procedures. The system was tested in the normal operating position with the system horizontally centered, no rotation, and with a height of 50" from mounting point to bottom of collimator.

**SEISMIC TEST PARAMETERS**

Building Code / Test Criteria	S <sub>DS</sub> (g)	z / h	I <sub>p</sub>	A <sub>FLX-H</sub> (g)	A <sub>RIG-H</sub> (g)	A <sub>FLX-V</sub> (g)	A <sub>RIG-V</sub> (g)
CBC 2022 / ICC-ES AC156	2.00	1.0	1.5	3.20	2.40	1.34	0.54

Note: The unit was full of contents during testing and remained functional before and after the ICC-ES AC156 test. The unit maintained structural integrity during and after the ICC-ES AC156 Test.

UUT<sub>z</sub>-2

**UNIT UNDER TEST (UUT)  
SUMMARY SHEET**



**Mounting Details:** Rigid floor mounted with 4 - 5/8" Gr. 8 bolts



<b>Manufacturer:</b> Siemens Healthcare GmbH	<b>Test Location:</b> Environmental Testing Laboratory
<b>Component:</b> Tilting Bucky Wall Stand	<b>Test Date:</b> July 2014
<b>Model Number:</b> 10681704	<b>Report Number:</b> SQ35-1415-02 Rev. 4
<b>UUT Function:</b> Wall stand for radiographic medical imaging	
<b>UUT Description:</b> Component of the YSIO and YSIO MAX Systems	

**UUT PROPERTIES**

Weight (lb)	Dimensions (inches)			Natural Frequency (Hz)		
	Width	Depth	Height	FB	SS	V
643	30	28.1	82.9	11.7	9.1	9.1

**SEISMIC TEST PARAMETERS**

Building Code / Test Criteria	S <sub>DS</sub> (g)	z / h	I <sub>P</sub>	A <sub>FLX-H</sub> (g)	A <sub>RIG-H</sub> (g)	A <sub>FLX-V</sub> (g)	A <sub>RIG-V</sub> (g)
CBC 2022 / ICC-ES AC156	2.00	1.0	1.5	3.20	2.40	1.34	0.54

Note: The unit was full of contents during testing and remained functional before and after the ICC-ES AC156 test. The unit maintained structural integrity during and after the ICC-ES AC156 Test.

UUT<sub>v</sub>-1

**UNIT UNDER TEST (UUT)  
SUMMARY SHEET**



**Mounting Details:** Rigid floor mounted with 4 - 3/8" bolts



BY: Mohammad Karim

DATE: 11/30/2022

**Manufacturer:** Siemens Healthcare GmbH **Test Location:** Environmental Testing Laboratory

**Component:** Polydoros RF RAD80 Generator **Test Date:** April 2013

**Model Number:** 10606040 **Report Number:** SQ35-1302-02 Rev. 1

**UUT Function:** Generator for radiography and fluoroscopy systems

**UUT Description:** Component of Multix Fusion system

**UUT PROPERTIES**

Weight (lb)	Dimensions (inches)			Natural Frequency (Hz)		
	Width	Depth	Height	FB	SS	V
424	22.5	50	21.4	31.9	32.6	12.3

**SEISMIC TEST PARAMETERS**

Building Code / Test Criteria	S <sub>DS</sub> (g)	z / h	I <sub>P</sub>	A <sub>FLX-H</sub> (g)	A <sub>RIG-H</sub> (g)	A <sub>FLX-V</sub> (g)	A <sub>RIG-V</sub> (g)
CBC 2022 / ICC-ES AC156	2.00	1.0	1.5	3.20	2.40	1.34	0.54

Note: The unit was full of contents during testing and remained functional before and after the ICC-ES AC156 test. The unit maintained structural integrity during and after the ICC-ES AC156 Test.

UUT<sub>v</sub>-2

**UNIT UNDER TEST (UUT)  
SUMMARY SHEET**



**Mounting Details:** Rigid wall mounted with 4 - 1/4" sheet metal screws



**Manufacturer:** Siemens Healthcare GmbH **Test Location:** Environmental Testing Laboratory

**Component:** Power Box **Test Date:** April 2013

**Model Number:** 10847706 **Report Number:** SQ35-1302-02 Rev. 1

**UUT Function:** Power box for wired detector

**UUT Description:** Component of Multix Fusion system

**UUT PROPERTIES**

Weight (lb)	Dimensions (inches)			Natural Frequency (Hz)		
	Width	Depth	Height	FB	SS	V
10	12.4	2.3	11	N/A	N/A	N/A

**SEISMIC TEST PARAMETERS**

Building Code / Test Criteria	S <sub>DS</sub> (g)	z / h	I <sub>P</sub>	A <sub>FLX-H</sub> (g)	A <sub>RIG-H</sub> (g)	A <sub>FLX-V</sub> (g)	A <sub>RIG-V</sub> (g)
CBC 2022 / ICC-ES AC156	2.00	1.0	1.5	3.20	2.40	1.34	0.54

Note: The unit was full of contents during testing and remained functional before and after the ICC-ES AC156 test. The unit maintained structural integrity during and after the ICC-ES AC156 Test.

UUT<sub>v</sub>-3

**UNIT UNDER TEST (UUT)  
SUMMARY SHEET**



**Mounting Details:** Rigid floor mounted with 4 - 3/8" bolts



**Manufacturer:** Siemens Healthcare GmbH | **Test Location:** Environmental Testing Laboratory

**Component:** Fluoroscopy Compact Imaging System | **Test Date:** April 2013

**Model Number:** 10569621 | **Report Number:** SQ35-1302-02 Rev. 1

**UUT Function:** Digital imaging solution for fluoroscopy and radiography

**UUT Description:** Component of Multix Fusion system

**UUT PROPERTIES**

Weight (lb)	Dimensions (inches)			Natural Frequency (Hz)		
	Width	Depth	Height	FB	SS	V
158	17.6	27.6	22.8	18.4	12.5	16.2

**SEISMIC TEST PARAMETERS**

Building Code / Test Criteria	S <sub>DS</sub> (g)	z / h	I <sub>P</sub>	A <sub>FLX-H</sub> (g)	A <sub>RIG-H</sub> (g)	A <sub>FLX-V</sub> (g)	A <sub>RIG-V</sub> (g)
CBC 2022 / ICC-ES AC156	2.00	1.0	1.5	3.20	2.40	1.34	0.54

Note: The unit was full of contents during testing and remained functional before and after the ICC-ES AC156 test. The unit maintained structural integrity during and after the ICC-ES AC156 Test.

UUT<sub>v</sub>-6

**UNIT UNDER TEST (UUT)  
SUMMARY SHEET**



**Mounting Details:** Rails and connecting parts of the component bolt with 2 -M10 bolts (20 bolts total) to unistrut grid spaced at 26.0" on center. The unistrut grid consisted of Unistrut P1001 rails anchored with 2 - 3/8" bolts at each intersection to the ceiling fixture framing spaced at 20" on center.



<b>Manufacturer:</b> Siemens Healthcare GmbH	<b>Test Location:</b> Environmental Testing Laboratory
<b>Component:</b> X-Ray Tube Stand w/ 3m Bridge	<b>Test Date:</b> April 2013
<b>Model Number:</b> 10746666 / 7042166	<b>Report Number:</b> SQ35-1302-02 Rev. 1
<b>UUT Function:</b> X-Ray stand, ceiling suspended, for use in radiography imaging, fully synchronized	
<b>UUT Description:</b> Component of Multix Fusion system	

**UUT PROPERTIES**

Weight w/ Rails (lb)	Dimensions (inches)			Natural Frequency (Hz)		
	Width	Depth	Height	FB	SS	V
750	167	119	33-104	N/A	N/A	N/A

The ceiling suspended tube stand moves laterally, rotates, and extends up and down to accommodate different patients and procedures. The system was tested in the normal operating position with the system horizontally centered, no rotation, and with a height of 72" from mounting point to bottom of collimator.

**SEISMIC TEST PARAMETERS**

Building Code / Test Criteria	S <sub>DS</sub> (g)	z / h	I <sub>p</sub>	A <sub>FLX-H</sub> (g)	A <sub>RIG-H</sub> (g)	A <sub>FLX-V</sub> (g)	A <sub>RIG-V</sub> (g)
CBC 2022 / ICC-ES AC156	2.00	1.0	1.5	3.20	2.40	1.34	0.54

Note: The unit was full of contents during testing and remained functional before and after the ICC-ES AC156 test. The unit maintained structural integrity during and after the ICC-ES AC156 Test.

UUT<sub>v</sub>-7

**UNIT UNDER TEST (UUT)  
SUMMARY SHEET**



**Mounting Details:** Rigid floor mounted with 4 - 5/8" bolts



**Manufacturer:** Siemens Healthcare GmbH | **Test Location:** Environmental Testing Laboratory

**Component:** Pro Non-Tilt Wall Stand | **Test Date:** August 2013

**Model Number:** 10681650 | **Report Number:** SQ35-1302-02 Rev. 1

**UUT Function:** Wall stand for radiographic medical imaging

**UUT Description:** Component of Multix Fusion system, included Varian Medical Systems portable digital detector (P/N: 10847569) installed in cassette tray during test. Cassette tray was modified to prevent detector from ejecting during testing.

**UUT PROPERTIES**

Weight (lb)	Dimensions (inches)			Natural Frequency (Hz)		
	Width	Depth	Height	FB	SS	V
452	30	16.6	82.9	12.8	14.5	13.6

**SEISMIC TEST PARAMETERS**

Building Code / Test Criteria	S <sub>DS</sub> (g)	z / h	I <sub>p</sub>	A <sub>FLX-H</sub> (g)	A <sub>RIG-H</sub> (g)	A <sub>FLX-V</sub> (g)	A <sub>RIG-V</sub> (g)
CBC 2022 / ICC-ES AC156	2.00	1.0	1.5	3.20	2.40	1.34	0.54

Note: The unit was full of contents during testing and remained functional before and after the ICC-ES AC156 test. The unit maintained structural integrity during and after the ICC-ES AC156 Test.

UUT<sub>v</sub>-8

**UNIT UNDER TEST (UUT)  
SUMMARY SHEET**



**Mounting Details:** Rigid floor mounted with 4 - 5/8" bolts



DATE: 11/30/2022

<b>Manufacturer:</b> Siemens Healthcare GmbH	<b>Test Location:</b> Environmental Testing Laboratory
<b>Component:</b> Top Tilting Wall Stand	<b>Test Date:</b> August 2013
<b>Model Number:</b> 10681670	<b>Report Number:</b> SQ35-1302-02 Rev. 1

**UUT Function:** Wall stand for radiographic medical imaging

**UUT Description:** Component of Multix Fusion system, included Varian Medical Systems portable digital detector (P/N: 10847569) installed in cassette tray during test. Cassette tray was modified to prevent detector from ejecting during testing.

**UUT PROPERTIES**

Weight (lb)	Dimensions (inches)			Natural Frequency (Hz)		
	Width	Depth	Height	FB	SS	V
490	30	28.1	82.9	8.4	9.1	9.3

**SEISMIC TEST PARAMETERS**

Building Code / Test Criteria	S <sub>DS</sub> (g)	z / h	I <sub>p</sub>	A <sub>FLX-H</sub> (g)	A <sub>RIG-H</sub> (g)	A <sub>FLX-V</sub> (g)	A <sub>RIG-V</sub> (g)
CBC 2022 / ICC-ES AC156	2.00	1.0	1.5	3.20	2.40	1.34	0.54

Note: The unit was full of contents during testing and remained functional before and after the ICC-ES AC156 test. The unit maintained structural integrity during and after the ICC-ES AC156 Test.

UUT<sub>v</sub>-9

**UNIT UNDER TEST (UUT)  
SUMMARY SHEET**



**Mounting Details:** Rigid floor mounted with 4 - 5/8" bolts



**Manufacturer:** Siemens Healthcare GmbH | **Test Location:** Environmental Testing Laboratory

**Component:** Patient Table | **Test Date:** August 2013

**Model Number:** 10273204 | **Report Number:** SQ35-1302-02 Rev. 1

**UUT Function:** Motorized patient table for support and positioning for image acquisition

**UUT Description:** Component of Multix Fusion system, included Varian Medical Systems portable digital detector (P/N: 10847569) installed in cassette tray during test. Cassette tray was modified to prevent detector from ejecting during testing.

**UUT PROPERTIES**

Weight (lb) with Patient	Dimensions (inches)			Natural Frequency (Hz)		
	Width	Depth	Height	FB	SS	V
1,546	31.5	94.8	20.3-37.6	2.4	27.4	7.4

The patient table moves laterally both ways and vertically to accommodate different patients and procedures. The system was tested in the normal operating position with the table horizontally centered, a table top height of 32", and a total simulated patient weight of 660lbs.

**SEISMIC TEST PARAMETERS**

Building Code / Test Criteria	S <sub>DS</sub> (g)	z / h	I <sub>p</sub>	A <sub>FLX-H</sub> (g)	A <sub>RIG-H</sub> (g)	A <sub>FLX-V</sub> (g)	A <sub>RIG-V</sub> (g)
CBC 2022 / ICC-ES AC156	2.00	1.0	1.5	3.20	2.40	1.34	0.54

Note: The unit was full of contents during testing and remained functional before and after the ICC-ES AC156 test. The unit maintained structural integrity during and after the ICC-ES AC156 Test.